



Urgent Medical Device Recall Notification

LMA[®] MAD Nasal[™] Intranasal Mucosal Atomization Device

October 27, 2016

To: Customer of Teleflex Medical Products

Teleflex Medical Incorporated ("Teleflex Medical") has issued a recall for the following product codes and lot numbers:

Product Code	Batch/ Lot#	Product Code	Batch/ Lot#	Product Code	Batch/ Lot#	
MAD100	160105	MAD130OS	160436	MAD300	160409	
	160137		160803		160422	
	160302	MAD140	160125		160432	
	160321		160218		160440	
	160402		160437		160500	
	160435		160610		160518	
	160506		160801		160602	
	160523		MAD140OS		160226	160611
	160609	160438			160621	
	160620	160727			160631	
	160707	160108			160701	
	MAD100OS	160802	MAD300		160117	160708
		160813			160126	160718
		160322			160145	160728
MAD110	160524	160146		160800		
	160630	160200		160804		
MAD110OS	160217	160219		160814		
	160507	160225		160816		
MAD130	160240	160231		160823		
	160312	160300		MAD300B 160410		
MAD130	160107	160313				
	160138	160327				
	160517	160400				

Teleflex Medical is recalling these products as they may not deliver a fully atomized plume of medication. Teleflex Medical has received complaints that the affected lots produced a straight stream instead of an atomized spray. The failure of the device to deliver an atomized plume may impair the effectiveness of the medication with which it is used. This can lead to serious injury or death in certain emergency situations, such as where the device is used in an off-label manner for needle-free delivery of drugs for reversal of life threatening narcotic overdose, reversal of life threatening hypoglycemia, or treatment of epileptic seizures.



Our records indicate that you have received products that are subject to this recall. We are now notifying our customers to take the following actions:

1. If you have affected stock, immediately discontinue use and quarantine any products with the catalog numbers listed above.
2. To return product, complete the enclosed Recall Acknowledgement Form and fax it to 1-855-419-8507, Attn: Customer Service or email to recalls@teleflex.com. This will allow us to document the amount of product you have on hand for return. A customer service representative will contact you with a Return Goods Authorization (RGA) Number and will provide instructions for the return of product to Teleflex Medical.
3. If you have no affected stock, please complete the enclosed Recall Acknowledgment Form and fax it to 1-855-419-8507, Attn: Customer Service or email to recalls@teleflex.com. This will allow us to document your receipt of this letter.

The U.S. Food and Drug Administration has been notified of this action.

Teleflex Medical is committed to providing high quality, safe and effective products. We sincerely apologize for any inconvenience this action may cause your operations. If you have any other questions, feel free to contact your local sales representative or Customer Service at 1-866-246-6990.

For and on behalf of Teleflex Medical,

Karen Boylan

Karen Boylan
VP, Global RA/QA

Enclosure